



KERALA GAZETTE

PUBLISHED BY AUTHORITY

Vol. LVI വാല്യം 56	THIRUVANANTHAPURAM, TUESDAY തിരുവനന്തപുരം, ചൊവ്വ	26th July 2011 2011 ജൂലൈ 26	No. 30 നമ്പർ
		4th Sravana 1933 1933 ശ്രാവണം 4	

PART III

Stores Purchase

Animal Husbandry Department

TENDER NOTICE

[No. I/2011-12]

No. 770/2007.

1st July 2011.

Sealed competitive tenders are invited from approved manufacturers or authorized distributors/dealers for supply of Double door Autoclave-1No. for the use of Biosafety Level-II facility at Avian Disease Diagnostic Laboratory, Thiruvalla.

Total amount— 1,79,859.

Earnest Money Deposit—1 % of the total quoted amount.

Cost of tender form:

Original— 400 + 12.5% VAT

Duplicate— 200 + 12.5% VAT

Last date and time upto which tender is issued—
25-7-2011, 3 p.m.

Last date and time of receiving tender—26-7-2011, 4 p.m.

Date and time of opening tender—27-7-2011, 2 p.m.

Address of office from which tender forms are available and tenders are received—Avian Disease

Diagnostic Laboratory, Manjadi P.O., Thiruvalla, Pin-689 105, Pathanamthitta Dist., Kerala.

All Kerala Government orders and laws applicable to tenders are applicable to this tender also. Full details of the above item will be available along with tender forms. The validity of this tender will be for 12 months from the date of this tender notification.

തിരുവല്ലാ പക്ഷിരോഗനിർണ്ണയശാലയിൽ ലബോറട്ടറിയിലെ ബയോസേഫ്റ്റി ലെവൽ സെക്കന്റ് സൗകര്യത്തിലേക്കായി ടബിൾ ഡോർ ഓട്ടോക്ലേവ്—1 എണ്ണം വിതരണം ചെയ്യുവാൻ താല്പര്യമുള്ള നിർമ്മാണ സ്ഥാപനങ്ങൾ/അംഗീകൃത വിതരണക്കാർ എന്നിവരിൽ നിന്നും മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിച്ചു കൊള്ളുന്നു.

അടങ്കൽ തുക— 1,79,859.

നിരതദ്രവ്യം—ആകെ തുകയുടെ 1%

ദർഘാസ് ഫോറത്തിന്റെ വില :

ഒറിജിനൽ— 400 + 12.5% മുല്യവർദ്ധിത നികുതി.

ഡ്യൂപ്ലിക്കേറ്റ്— 200 + 12.5% മുല്യവർദ്ധിത നികുതി.

ദർഘാസ് ഫോറം നൽകുന്ന അവസാന തീയതിയും സമയവും—25-7-2011, 3 മണി.

ദർഘാസ് ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതിയും സമയവും—26-7-2011, 4 മണി.

ദർഘാസ് തുറന്ന് പരിശോധിക്കുന്ന തീയതിയും സമയവും—27-7-2011, 2 മണി.

ദർഘാസ് ഫോറം നൽകുകയും സ്വീകരിക്കുകയും ചെയ്യുന്ന ഓഫീസിന്റെ വിലാസം — പക്ഷിരോഗനിർണ്ണയശാല, മഞ്ഞാടി പി.ഒ., തിരുവല്ല, പത്തനംതിട്ട ജില്ല, കേരള, പിൻ-689 105.

കേരള സർക്കാർ ദർഘാസുകൾക്ക് ബാധകമാകുന്ന എല്ലാ നിയമങ്ങളും ഈ ദർഘാസിനും ബാധകമാണ്. ഉപകരണങ്ങളുടെ എല്ലാ വിവരങ്ങളും ദർഘാസ് ഫോറത്തിനൊപ്പം ലഭ്യമാണ്. ദർഘാസിന്റെ കാലാവധി ഈ ദർഘാസ് പരസ്യപ്പെടുത്തിയതു മുതൽ 12 മാസത്തേക്കായിരിക്കുന്നതാണ്.

SPECIFICATIONS

Double Door Autoclave

Chamber volume— 35 Liter

Chamber dimension — 300 × 500 mm.

Loading capacity —250 ml. flasks

10 air changes per hour

Fully automated operation with digital control and high safety lid

Microprocessor based controller with display for temperature and time

Heavy duty jacket

Construction chamber —SS 304

Operating pressure —15 PSI

Operating temperature —121 degree celcius

SPECIAL TERMS AND CONDITIONS

1. The offer should be for supply from ready stock in original packing of the manufacturer and the tenderers should give the name of the manufacturer of the item.

2. Tenderers should submit a crossed demand draft for 1% of the quoted amount drawn in favour of Deputy Director, Avian Disease Diagnostic Laboratory, Thiruvalla along with the tender as Earnest Money Deposit.

3. Rate quoted should be in Indian Currency and or Foreign Currency on CIP basis.

4. If the supply order is placed in Indian currency, FOR delivery to destination and if in foreign currency, the delivery should be as per the terms and conditions of Letter of Credit.

5. Ex godown rate and free on rail despatch at stations offered by firms are not acceptable and such tenders will be rejected.

6. The firm who are not original manufacturers should produce the authorization certificate by the manufacturer.

7. Sufficient brochures and technical write ups of the quoted items should be provided along with the tender.

8. A minimum of one year guarantee should be there for every tendered item.

9. Tenderers should submit a preliminary agreement in the Kerala Stamp Paper worth 100 along with the tender. In the absence of such agreement, tender will be rejected.

10. The manufacturer should have ISO certification or other valid accreditation.

11. Annual Maintenance Contract for 5 years should be there with the company.

12. Supply of the items should be affected within 15 days on receipt of supply order. No extension of time will be allowed.

13. A minimum of three demonstrations should be done by the firm/authorized technical hands after installation.

14. The Earnest Money Deposit furnished by the firms as demand draft will be released after satisfactory completion of contracts and the tenderers should collect the same from this office.

15. Payments will be made within 15 days from the date of acceptance of the item by the Deputy Director, Avian Disease Diagnostic Laboratory, Thiruvalla after satisfactory installation and demonstration of the item.

Avian Disease
Diagnostic Laboratory,
Manjady, Thiruvalla.

(Sd.)
Deputy Director.